



UNITED STATES DEPARTMENT OF COMMERCE Patent and Trademark Office

Address: COMMISSIONER OF PATENTS AND TRADEMARKS Washington, D.C. 20231

DATE MAILED:

			· ·				
	APPLICATION NO. FILING DATE		FIRST NAMED INVENTOR			ATTORNEY DOCKET NO.	
	09/556,72	6 04/21/00	THOMPSON		P	760-24 DIV 2	
Г	-		QM12/1003	¬ [EXAMINER		
	HOFFMANN & BARON LLP		Section and the the the site	_	PREBILIC, P		
		CHO TURNPIKE			ART UNIT	PAPER NUMBER	
	SYOSSET N	Y 11/ J 1			3738	#3	

Please find below and/or attached an Office communication concerning this application or proceeding.

Commissioner of Patents and Trademarks

10/03/01

Office Action Summary

Application No. 09/556,726 Applicate(s)

Examiner

Paul Prebilic

Art Unit 3738

Thompson



-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --**Period for Reply** A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136 (a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). - Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). Status 1) 🔀 Responsive to communication(s) filed on Jul 24, 2000 2b) X This action is non-final. 2a) 🗔 This action is **FINAL**. 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11; 453 O.G. 213. **Disposition of Claims** 4) X Claim(s) 19, 20, and 22-26 is/are pending in the application. 4a) Of the above, claim(s) ______ is/are withdrawn from consideration. Claim(s) is/are allowed. Claim(s) 19 and 20 is/are rejected. Claim(s) 22-26 is/are objected to. Claims _____ are subject to restriction and/or election requirement. **Application Papers** 9) The specification is objected to by the Examiner. The drawing(s) filed on is/are objected to by the Examiner. The proposed drawing correction filed on is: a) approved b) disapproved. 12) 📙 The oath or declaration is objected to by the Examiner. Priority under 35 U.S.C. § 119 13) Acknowledgement is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d). a) \square All b) \square Some* c) \square None of: 1. Certified copies of the priority documents have been received. Certified copies of the priority documents have been received in Application No. _______. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). *See the attached detailed Office action for a list of the certified copies not received. 14) Acknowledgement is made of a claim for domestic priority under 35 U.S.C. § 119(e). Attachment(s) 15) X Notice of References Cited (PTO-892) 18) Interview Summary (PTO-413) Paper No(s). 16) Notice of Draftsperson's Patent Drawing Review (PTO-948) 19) Notice of Informal Patent Application (PTO-152) 17) X Information Disclosure Statement(s) (PTO-1449) Paper No(s). _______ 20) Other:

Art Unit: 3738

Specification

The disclosure is objected to because of the following informalities:

The continuing data on page 1 of the specification needs to be updated with the current status of each parent application.

On page 3, line 14, it appears that "14" should be --140-- based upon original claim 5 which give original support thereto.

Appropriate correction is required.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless --

- (a) the invention was known or used by others in this country, or patented or described in a printed publication in this or a foreign country, before the invention thereof by the applicant for a patent.
- (e) the invention was described in a patent granted on an application for patent by another filed in the United States before the invention thereof by the applicant for patent, or on an international application by another who has fulfilled the requirements of paragraphs (1), (2), and (4) of section 371(c) of this title before the invention thereof by the applicant for patent.

Claim 19 is rejected under 35 U.S.C. 102(a) as being anticipated by Myers et al (WO 95/05132). Myers et al discloses a braided stent (see Figure 9) intimately bound (see Figure 8) to a biaxially expanded PTFE graft; see the whole document. For this reason, the claim language is fully met because the stent and graft would inherently expand and contract together such that the expansion ratios thereof would be virtually the same.

Application/Control Number: 09/556,726

Art Unit: 3738

Claim 19 is rejected under 35 U.S.C. 102(e) as being anticipated by Myers et al (US 5,700285). Myers et al discloses a braided stent (see Figure 9) intimately bound (see Figure 8) to a biaxially expanded PTFE graft; see the whole document. For this reason, the claim language is fully met because the stent and graft would inherently expand and contract together such that the expansion ratios thereof would be virtually the same; in other words, less than 25% as claimed.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claim 20 is rejected under 35 U.S.C. 103(a) as being unpatentable over Banas et al (US 5,749,880) in view of Wallsten (US 4,954,126). Banas et al discloses many stent types (see column 12, line 49 et seq.) which can be bonded to uniaxially expanded PTFE with no nodules (only nodes, see Figure 18A to 19B). However, Banas fails to specifically disclose using a braided stent as claimed. Wallsten, however, teaches that it was known to the art to make braided self-expanding stents. Hence, it is the Examiner's position that it would have been obvious to use a braided self-expanding stent, as taught by Wallsten, for the same reasons that Wallsten desires the same and so that self-expansion and even distribution of pressure can be brought to the Banas et al device.

Art Unit: 3738

Allowable Subject Matter

Claims 22-26 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Conclusion

Applicant should specifically point out the support for any amendments made to the disclosure, including the claims (MPEP 714.02 and 2163.06). Due to the procedure outlined in MPEP 2163.06 for interpreting claims, it is noted that other art may be applicable under 35 USC 102 of 35 USC 103(a) once the aforementioned problem is corrected.

Applicant is respectfully requested to provide a list of all copending applications that set forth similar subject matter to the present claims. A copy of such copending claims is respectfully requested in response to this Office action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Paul Prebilic whose telephone number is (703) 308-2905. The examiner normally be reached on Monday-Thursday from 6:30 AM to 5:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott, can be reached on (703) 308-2111. The fax phone number for this Technology Center is (703) 872-9301.

Any inquiry of a general nature or relating to the status of this application should be directed to the Technology Center 3700 receptionist whose telephone number is (703) 308-0858.

Paul Prebilic

Primary Examiner

Paul Preliet

Art Unit 3738

Attachment for PTO-948 (Rev. 03/01, or earlier) 6/18/01

The below text replaces the pre-printed text under the heading, "Information on How to Effect Drawing Changes," on the back of the PTO-948 (Rev. 03/01, or earlier) form.

INFORMATION ON HOW TO EFFECT DRAWING CHANGES

1. Correction of Informalities -- 37 CFR 1.85

New corrected drawings must be filed with the changes incorporated therein Identifying indicia, if provided, should include the title of the invention, inventor's name, and application number, or docket number (if any) if an application number has not been assigned to the application. If this information is provided, it must be placed on the front of each sheet and centered within the top margin. If corrected drawings are required in a Notice of Allowability (PTOL-37), the new drawings MUST be filed within the THREE MONTH shortened statutory period set for reply in the Notice of Allowability. Extensions of time may NOT be obtained under the provisions of 37 CFR 1 136(a) or (b) for filing the corrected drawings after the mailing of a Notice of Allowability. The drawings should be filed as a separate paper with a transmittal letter addressed to the Official Draftsperson.

2. Corrections other than Informalities Noted by Draftsperson on form PTO-948.

All changes to the drawings, other than informalities noted by the Draftsperson. MUST be made in the same manner as above except that, normally, a highlighted (preferably red ink) sketch of the changes to be incorporated into the new drawings MUST be approved by the examiner before the application will be allowed. No changes will be permitted to be made, other than correction of informalities, unless the examiner has approved the proposed changes.

Timing of Corrections

Applicant is required to submit the drawing corrections within the time period set in the attached Office communication. See 37 CFR 1.85(a).

Failure to take corrective action within the set period will result in **ABANDONMENT** of the application